

AirSim Bronchi

User Manual



Product Code: AA51100

Size: 46 x 26 x 33 cm

Ship Weight: 3.3 kg

Package contents:

- 1 AirSim Bronchi (assembled and ready to use)
- 1 Black carrier case
- 1 USB pen drive user manual
- 1 Bottle of lubrication
- 1 Adult Bronchi Lung Bag set

Before you begin....

- Remove the protective head foam
- Secure the manikin to a suitable surface by adjusting the screw on the suction cups (the surface needs to be non porous to ensure sufficient suction)
- Loosen the black neck lock to provide easier head movement. Ensure the head is in the desired position and then tighten the neck lock again
- Check the amount of air in the tongue is sufficient, it can be inflated or deflated by a syringe to adjust the difficulty during intubation
- Check the lung bags and stomach bag are attached to the airway and oesophagus
- Spray generous amounts of lubrication on the airway and equipment
- You are now ready to use the manikin

The AirSim Bronchi is ideal for training a range of techniques including:

- Diagnostic Bronchoscopy
- Lung isolation techniques using left and right endo-bronchial tubes and bronchial blockers
- Lung suctioning techniques
- Single naso-tracheal intubation
- Bag and Mask ventilation techniques
- Full range of supraglottic devices
- Direct laryngoscopy
- Endotracheal tube insertion
- Awake fibre optic examination
- Single lung isolation techniques
- Combi tube insertion

We recommend the following equipment sizes for optimal performance:

- 7.0-7.5 mm ID for nasal intubation
- 8.0-9.0 mm ID for oral intubation
- Size 3-5 for LMA laryngeal masks
- Size 35F-37F for endo-bronchial tubes
- Similar respective sizes for other supraglottic devices

Care and Maintenance

Store in clean, dry conditions away from heat and direct sunlight; avoid contact with metals, solvents, oils or greases and strong detergents. When the product is not in use please store in the black carrier case provided.

Thoroughly wash the AirSim airway in warm water. Please use warm soapy water or something similar until all visible foreign matter and residue is removed.

Mild detergents or enzymatic cleaning agents may be used on the airway in accordance with the manufacturer's instructions and at the proper dilution. The detergent must not contain skin or mucous membrane irritants.

Please do not use any of the following when cleaning the AirSim product range

- Germicides, disinfectants, or chemical agents such as glutaraldehyde (e.g. Cidex®),
- Ethylene oxide, phenol-based cleaners or iodine-containing cleaners

Warranty

TruCorp warrants this unit to be free of defects in materials and workmanship and to give satisfactory service for a period of 1 year from the date of delivery.

This ensures that our customers receive maximum coverage on each product. If the unit should malfunction it must be returned to the factory for evaluation. Upon examination by TruCorp, if the unit is found to be defective it will be repaired or replaced at no charge. However this warranty is VOID, if; the unit shows evidence of having been tampered with or shows evidence of having been damaged by excessive heat, the use of sharp instruments, misapplication, misuse or other operating conditions outside of TruCorp's control. Components which wear or are damaged by misuse are not warranted and will be charged for if repair has been approved.

Please direct all warranty and repair requests/inquires to:



The Mount Business Park
2 Woodstock Link,
Belfast, BT6 8DD

info@trucorp.com / rcolhoun@trucorp.com



Tel: +44 (0) 28 9073 7281 Fax: +44 (0) 28 9073 7282